



technoclone

ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQRR00AZ

TECHNOZYM[®] ADAMTS-13 Antigen ELISA Kit 96T.REF 5450601
LOT RR01C00.02
🕒 2021-05-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x 12	Adamts13 Ag Plate	RRS92B0	2021-11-30
1 x 0.3mL	Adamts13 Ag Conjugate	RQV91B0	2021-05-31
1x 80mL	Wash buffer Concentrate	RNA93B0	2021-09-30
1 x 90mL	Incubation buffer	RNB93B0	2021-10-31
1 x 12mL	TMB Substrate	RKN92B0	2021-07-31
1 x 12mL	Stop solution	RKK92B0	2021-06-30
1 x 0.5mL	Adamts13 Ag Calibrator 1	1RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 2	2RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 3	3RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 4	4RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Calibrator 5	5RT91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag High Control	RRY91B0	2021-08-31
1 x 0.5mL	Adamts13 Ag Low Control	RRZ91B0	2021-08-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Δ Cal 1 – Cal 5	≥ 0.7	0.83	OK
Verwendbarkeit Usability	High Control (RRY91B0)	0.47 – 0.79 IU/mL	0.62	OK
	Low Control (RRZ91B0)	0.12 – 0.28 IU/mL	0.22	OK

Durchführung von
Prepared byMichael KAFKA 21. JAN. 2020

Datum / Date:

Freigabe von
Approved byFikret NASUFI 21. JAN. 2020

Datum / Date: