

CERTIFICATE OF ANALYSIS

CHROMOGENIX

**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME:	<u>FXa, 10 x 71 nkat</u>	EXPIRATION DATE:	<u>2022-04</u>
PRODUCT NUMBER:	<u>82098539</u>		
LOT NUMBER:	<u>N1199786</u>		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			PASS/FAIL
FACTOR Xa, 71 nkat (650515)					
ACTIVITY	61	81	nkat/vial	69	PASS
STANDARD CURVE (r ²)	0.990	1.000	--	1.000	PASS
CONTROL RECOVERY	66.0	73.0	nkat/vial	69.3	PASS

PREPARED BY: BHUMI VASOYA
PRINT

Bhumi Vasoya
SIGNATURE

DATE: 11/13/2019

TITLE: QC ANALYST-I

APPROVAL SIGNATURE: DEBRA HARZ
PRINT

Debra Harz
SIGNATURE

DATE: 11/13/19

TITLE: QC TECH SUPERVISOR