

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2403, 25 mg Substrate	EXPIRATION DATE:	2023-01
PRODUCT NUMBER:	82225439		
LOT NUMBER:	N0101536		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	24.9	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.1	PASS

PREPARED BY: Maia Oettinger *Maia Oettinger* DATE: 2-12-2020
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Hays *Debra Lee Hays* DATE: 2-12-2020
 PRINT SIGNATURE

TITLE: QC Technical Supervisor