



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0U00.02

AK-Calibrant 4 x 1 mL

REF 5010004
LOT 0U83B00.07
⌚ 2021-09-30

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	AK Calibrant A	1N84B00	2021-09-30
1x1mL	AK Calibrant B	3L88B00	2021-09-30
1x1mL	AK Calibrant C	3M8DB00	2021-10-31
1x1mL	AK Calibrant D	3N8EB00	2021-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P92B00 Technoplastin HIS	94.7 – 119.3 % 0.84 – 1.06 INR	103.9 0.97	OK OK
	CCA LOT 3P91B00 Technoplastin HIS	23.0 – 29.0 % 2.78 – 3.50 INR	26.3 3.13	OK OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Eva Maria HAGER

25. FEB. 2020

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

26. FEB. 2020

Datum / Date: