

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
A Werfen Company

PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE: 2021-10
PRODUCT NUMBER :	82258563	
LOT NUMBER:	N1199920	

TEST DESCRIPTION	SPECIFICATIONS			RESULTS	PASS/FAIL
	MIN	MAX	UNITS		
HIGH RANGE					
Slope	0.33	0.57	--	0.40	PASS
Intercept	---	0.12	--	0.12	PASS
r	0.990	---	--	0.993	PASS
LOW RANGE					
Slope	2.8	6.9	--	4.3	PASS
Intercept	---	0.21	--	0.18	PASS
r	0.990	---	--	0.999	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N1199918	3 mL
COAMATIC FVIII S-2765 + I-2581	N1199919	6 mL
COAMATIC FVIII BUFFER	N1099583	24 mL

· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Nehal Patel *N.M. Patel* DATE: 12/13/2019 *12-13-19*
 PRINT SIGNATURE

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 12/13/2019 *12-13-19*
 PRINT SIGNATURE

TITLE: Technical Supervisor