



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3400.01

TECHNOVIEW® Rivaroxaban Control Low 5x1 mL

REF 5090172

LOT 3401C00.01

🕒 2022-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.2 – 7.9	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		40.3 – 74.8 ng/mL	57.5	OK
Verwendbarkeit Usability Ceveron Alpha®	TECHNOCHROM® anti-Xa	40.3 – 74.8 ng/mL	49.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

24. APR. 2020
Datum / Date:

Freigabe von
Approved by


Eva Maria HAGER

24. APR. 2020
Datum / Date: