



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQRP00AZ

## TECHNOZYM® ADAMTS-13 INH ELISA Kit 48T

REF 5450451

LOT RP05C0S.02

🕒 2021-10-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1 x 6	Adamts13 INH Plate	RPS9HBS	2021-12-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV92B0	2021-11-30
1x 80mL	Wash buffer Concentrate	RNA01C0	2022-02-28
1 x 90mL	Incubation buffer	RNB93B0	2021-10-31
1 x 12mL	TMB Substrate	RKN01C0	2022-01-31
1 x 12mL	Stop solution	RKK01C0	2022-01-31
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH High Control	RPY92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Low Control	RPZ92B0	2021-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.69	OK
	Calibrator 5	< 0.1	0.06	OK
Verwendbarkeit Usability	High Control (RPY92B0)	56.36 - 93.94 U/mL	80.03	OK
	Low Control (RPZ92B0)	15.11 - 35.27 U/mL	26.22	OK

Durchführung von  
Prepared by

Michael KAFKA 

11. MAI 2020

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI 

12. MAI 2020

Datum / Date: