

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ7PT00AZ
---	---	-----------

**TECHNOZYM® ADAMTS-13 INH Control Set 2x0.5mL**

REF      5450463  
LOT      7PT01C0.01  
      2021-11-30

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1 x 0.5mL	Adamts13 INH High Control	RPY92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Low Control	RPZ92B0	2021-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	High Control (RPY92B0)	56.36 – 93.94 U/mL	74.40	OK
	Low Control (RPZ92B0)	15.11 – 35.27 U/mL	27.85	OK

**Durchführung von**  
**Prepared by**

Fikret NASUFI *Fikret Nasufi*

09. APR. 2020  
Datum / Date:

**Freigabe von**  
**Approved by**

Michael KAFKA *[Signature]*

09. APR. 2020  
Datum / Date: