



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ2600.01

TECHNOVIEW® Arixtra CAL 5x1 mL

REF 5090010

LOT 2601C00.01

🕒 2022-05-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x1 mL	TECHNOVIEW Arixtra Cal 1	2181B01	2022-05-31
1x1 mL	TECHNOVIEW Arixtra Cal 2	2281B01	2022-05-31
1x1 mL	TECHNOVIEW Arixtra Cal 3	2381B01	2022-05-31
1x1 mL	TECHNOVIEW Arixtra Cal 4	2481B01	2022-05-31
1x1 mL	TECHNOVIEW Arixtra Cal 5	2581B01	2022-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Arixtra Cont low LOT 3U81B01	0.30 – 0.56 µg/mL	0.49	OK
	TECHNOVIEW Arixtra Cont high LOT 3V81B01	0.80 – 1.34 µg/mL	1.24	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
Stefan STEINER

09. JUNI 2020

Datum / Date:

Freigabe von  
Approved by

  
Christoph ENNE

09. JUNI 2020

Datum / Date: