



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3V00.01

TECHNOVIEW Arixtra CONT H 6x1 mL

REF 5090014

LOT 3V01C00.01

 2022-03-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa Arixtra	1.00 – 1.66 µg/mL	1.33	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa Arixtra	1.00 – 1.66 µg/mL	1.28	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

15. APR. 2020

Datum / Date:

Freigabe von
Approved by


Eva Maria HAGER

15. APR. 2020

Datum / Date: