
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00.02
---	---	-----------

TECHNOCHROM[®] FXIII 3 x 3 mL

REF	5360010
LOT	0Z02C00.01
	2021-04-30

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	2002100/1	2021-04-30
3 x 3 mL	FXIII Detection Reagent	2002100/2	2021-04-30
3 x 3 mL	FXIII NADPH	2002100/3	2021-04-30
3 x 1 mL	FXIII Inhibitor Reagent	2002100/4	2021-04-30
1 x 6 mL	Stabilizer solution	2002100/5	2021-04-30

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P02C00	60.8 – 101.3 % 74.1
	Coagulation Control A Lot: 3P91B00	24.0 – 40.0 % 26.0

Durchführung von
Prepared by

Michael KAFKA 

25. MAI 2020
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

25. MAI 2020
Datum / Date: