


	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRB0.03 Att.07 RB05C00.01
---	---	----------------------------------

TECHNOZYM vWF:CBA Collagen Typ III ELISA Kit 96T.

REF	5450301
LOT	RB05C00.01
	2022-01-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	vWF Calibrator 1	1ND01C0	2022-02-28
1x	0.5 mL	vWF Calibrator 2	2ND01C0	2022-02-28
1x	0.5 mL	vWF Calibrator 3	3ND01C0	2022-02-28
1x	0.5 mL	vWF Calibrator 4	4ND01C0	2022-02-28
1x	0.5 mL	vWF Calibrator 5	5ND01C0	2022-02-28
1x	0.5 mL	vWF High Control	AND01C0	2022-02-28
1x	0.5 mL	vWF Low Control	BND01C0	2022-02-28
1x	12 mL	TMB Substrate	RKN01C0	2022-01-31
1x	12 mL	Stop solution	RKK01C0	2022-01-31
1x	80 mL	Wash buffer concentrate	RNA03C0	2022-05-31
1x	90 mL	Incubation buffer	RNB01C0	2022-02-28
1x	-	vWF:CBA Plate Collagen Typ III	RNC03C0	2022-05-31
1x	0.3 mL	CBA Conjugate	RNE01C0	2022-02-28

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD CaI1-CaI5	≥ 0.7	1.4	OK
Kontrollwiederfindung Recovery of controls	vWF High Control LOT AND01C0	0.78 - 1.30 IU/mL	1.21 IU/mL	OK
	vWF Low Control LOT BND01C0	0.20 - 0.48 IU/mL	0.39 IU/mL	OK

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRB0.03 Att.07 RB05C00.01
---	---	----------------------------------

Durchgeföhrt von
Prepared by

Michael KAFKA

24. JUNI 2020

Datum / Date:

Freigabe von
Approved by

Christoph ENNE

24. JUNI 2020

Datum / Date: