


	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQR100.01 Att.05 R102C00.01
---	---	-----------------------------------

TECHNOZYM vWF:CBA Collagen Typ VI ELISA Kit 96T.

REF	5450321
LOT	R102C00.01
	2021-05-31

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5 mL	vWF Calibrator 1	1ND91B0	2021-05-31
1x	0.5 mL	vWF Calibrator 2	2ND91B0	2021-05-31
1x	0.5 mL	vWF Calibrator 3	3ND91B0	2021-05-31
1x	0.5 mL	vWF Calibrator 4	4ND91B0	2021-05-31
1x	0.5 mL	vWF Calibrator 5	5ND92B0	2021-06-30
1x	0.5 mL	vWF High Control	AND91B0	2021-05-31
1x	0.5 mL	vWF Low Control	BND91B0	2021-05-31
1x	12 mL	TMB Substrate	RKN92B0	2021-07-31
1x	12 mL	Stop solution	RKK92B0	2021-06-30
1x	80 mL	Wash buffer concentrate	RNA93B0	2021-09-30
1x	90 mL	Incubation buffer	RNB93B0	2021-10-31
1x	-	vWF:CBA Plate Collagen Typ VI	RNN61B2	2022-01-31
1x	0.3 mL	CBA Conjugate	RNH81B1	2021-06-30

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Batch table	present	confirmed	OK
Verwendbarkeit Usability	deltaOD Cal1-Cal5	≥ 0.7	2.5	OK
Kontrollwiederfindung Recovery of controls	vWF High Control LOT AND91B0	0.94 - 1.56 IU/mL	1.34 IU/mL	OK
	vWF Low Control LOT BND91B0	0.22 - 0.52 IU/mL	0.38 IU/mL	OK

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQR100.01 Att.05 R102C00.01
---	---	-----------------------------------

Durchgeföhrt von
Prepared by

Michael KAFKA 

28. FEB. 2020

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

28. FEB. 2020

Datum / Date: