

**CERTIFICATE OF ANALYSIS**

**CHROMOGENIX**

**Instrumentation  
Laboratory**  
A Werfen Company

PRODUCT NAME:	<u>FXa, 10 x 71 nkat</u>	EXPIRATION DATE:	<u>2022-11</u>
PRODUCT NUMBER:	<u>82098539</u>		
LOT NUMBER:	<u>N0504136</u>		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
<b>FACTOR Xa, 71 nkat (650515)</b>					
ACTIVITY	61	81	nkat/vial	77	PASS
STANDARD CURVE (r <sup>2</sup> )	0.990	1.000	--	0.999	PASS
CONTROL RECOVERY	66.0	73.0	nkat/vial	71.2	PASS

PREPARED BY: Maria Oettinger *Maria Oettinger* DATE: 5-18-20  
 PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 5-18-20  
 PRINT SIGNATURE

TITLE: QC Technical Supervisor