CERTIFICATE OF ANALYSIS

CHROMOGENIX Instrumentation Laboratory A Werfen Company

PRODUCT NAME:	COATEST APC Resistance Control Plasma Level 2	EXPIRATION DATE: 2022-03
PRODUCT NUMBER:	82266863	
LOT NUMBER:	N0303159	

TEST DESCRIPTION		SPECIFICATIONS			RESULTS	DISPOSITION
Analysis with COATEST APC R:		**MIN**	**MAX**	**UNITS**		**PASS/FAIL**
APT Time	$CaCl_2$	27	44	seconds	44	PASS
	APC/$CaCl_2$	--	101	seconds	97	PASS
APC Ratio		--	2.3	--	2.2	PASS
Analysis with COATEST APC R V:						
APT Time	$CaCl_2$	27	40	seconds	35	PASS
APC Ratio		1.4	1.9	--	1.8	PASS

We hereby certify that the above product(s):
· have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all conditions set forth with these documents;
· meet all applicable requirements of the 98/79/EC Directive according to which they are CE marked.

PREPARED BY: Maria Oettinger DATE: 4-30-2020
PRINT SIGNATURE

TITLE: QC Analyst III

APPROVAL SIGNATURE: Debra Lee Harz DATE: 4/30/2020
PRINT SIGNATURE

TITLE: QC Technical Supervisor

526 Route 303, Orangeburg, New York 10962, Telephone 845-365-8000