



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQRXXAZ

TECHNOZYM® t-PA Combi Actibind ELISA Kit 96T

REF TC16000  
LOT 160002C0.01  
🕒 2021-10-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x	anti t-PA Plate	RGN01C0	2022-04-30
1x0.3mL	POX anti t-PA	RKM01C0	2022-04-30
1x80mL	Wash buffer Concentrate	RNA01C0	2022-02-28
1x90mL	Incubation buffer	RNB01C0	2022-02-28
1x12mL	TMB Substrat	RKN01C0	2022-01-31
1x12mL	Stop solution	RKK01C0	2022-01-31
1x0.5mL	Standard t-PA	RAW91B0	2021-10-31
1x20mL	Detection mix buffer- TRIS	RDA01C0	2022-05-31
1x	Detection mixture	RBA71B1	2022-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1 activity	> 0.7	1.07	OK
	Calibrator 5 activity	< 0.1	0.04	OK
	Calibrator 1 antigen	> 1.0	1.37	OK
	Calibrator 5 antigen	< 0.1	0.00	OK

Durchführung von  
Prepared by

Michael KAFKA

25. JUNI 2020

Datum / Date:

Freigabe von  
Approved by

Fikret NASUFI

25. JUNI 2020

Datum / Date: