

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQRE00AZ
---	---	----------

## TECHNOZYM<sup>®</sup> PAI-1 Actibind ELISA Kit 96T

REF TC16075  
LOT RE02C00.01  
 2021-07-31

Inhalt/ Content	Reagenz/Reagent	LOT	
1 x	anti t-PA coated plate Actibind	RGP02C0	2022-03-31
1 x 0.3 mL	anti PAI-1 Conjugate	RKQ02C0	2022-04-30
1 x 80 mL	Wash buffer Concentrate	RNA01C0	2022-02-28
1 x 90 mL	Incubation buffer	RNB01C0	2022-02-28
1 x 12 mL	TMB Substrat	RKN01C0	2022-01-31
1 x 12 mL	Stop solution	RKK01C0	2022-01-31
1 x 0.2 mL	Calibrator 1 active PAI-1	1DW95B0	2021-09-30
1 x 0.2 mL	Calibrator 2 active PAI-1	2DW92B0	2021-07-31
1 x 0.2 mL	Calibrator 3 active PAI-1	3DW92B0	2021-07-31
1 x 0.2 mL	Calibrator 4 active PAI-1	4DW92B0	2021-07-31
1 x 0.2 mL	Calibrator 5 active PAI-1	5DW91B0	2021-07-31
1 x 0.2 mL	High Control active PAI-1	ADW91B0	2021-07-31
1 x 0.2 mL	Low Control active PAI-1	BDW91B1	2021-07-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	Top OD Cal 1	1 – 2.5	1.35	OK
	OD Cal 5	< 0.2	0.06	OK
	High Control (ADW91B0)	12.80 – 21.34 IU/mL	18.97	OK
	Low Control (BDW91B1)	1.39 – 4.16 IU/mL	2.44	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasufi*

24. APR. 2020

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *[Signature]*

24. APR. 2020

Datum / Date: