

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2238, 25 mg Substrate	EXPIRATION DATE: 2022-12
PRODUCT NUMBER:	82032439	
LOT NUMBER:	N1290409	

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.6	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.3	PASS

PREPARED BY: BHUMI VASOYA
PRINT

Bhumi Vasoya
SIGNATURE

DATE: 12-12-19

TITLE: QC ANALYST - I

APPROVAL SIGNATURE: DEBRA HARZ
PRINT

Debra Harz
SIGNATURE

DATE: 01/02/20

TITLE: QC TECH SUPERVISOR