

CERTIFICATE OF ANALYSIS

CHROMOGENIX



**Instrumentation
Laboratory**

A Werfen Company

PRODUCT NAME:	S-2765, 25 mg Substrate	EXPIRATION DATE:	2022-12
PRODUCT NUMBER:	82141339		
LOT:	N1290406		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
SUBSTRATE CONCENTRATION	23.7	26.3	mg/vial	25.5	PASS
FREE p-NITROANILINE	0	0.5	mol %	0.3	PASS

PREPARED BY: BHUMI VASOYA Bhumi Vasoya DATE: 12-12-2019
 PRINT SIGNATURE

TITLE: QC ANALYST - 1

APPROVAL SIGNATURE: DEBRA HARZ Debra Harz DATE: 12-18-19
 PRINT SIGNATURE

TITLE: QC TECH SUPERVISOR