
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ1P00.01
---	---	-----------

### Coagulation Control N 5x1 mL

REF      5020040  
LOT      1P02C00.02  
      2021-07-31

<b>Analytische Prozedur Analytical Procedure</b>		<b>Anforderung Requirement</b>	<b>Ergebnis Result</b>	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Ceveron® alpha	Siron LS	21.1 – 28.5 sec	26.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
Eva Maria HAGER

14. JULI 2020  
Datum / Date:

**Freigabe von  
Approved by**

  
Suzana GRESAK PODAKAR

14. JULI 2020  
Datum / Date: