



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0Y00.01

TECHNOVIEW UFH CAL 5x1 mL

REF 5090070

LOT 0Y01C00.01

⌚ 2022-05-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1mL	TECHNOVIEW UFH Cal 1	5F91B01	2022-05-31
1x1mL	TECHNOVIEW UFH Cal 2	5G91B01	2022-05-31
1x1mL	TECHNOVIEW UFH Cal 3	5H91B01	2022-05-31
1x1mL	TECHNOVIEW UFH Cal 4	5J91B01	2022-05-31
1x1mL	TECHNOVIEW UFH Cal 5	5K91B01	2022-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW UFH Cont low LOT 3C81B01	0.11 – 0.20 IU/mL	0.19	OK
	TECHNOVIEW UFH Cont high LOT 3J81B01	0.47 – 0.79 IU/mL	0.66	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Stefan STEINER

09. JUNI 2020

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

09. JUNI 2020

Datum / Date: