
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ0600.01
---	---	-----------

## TECHNOVIEW® Dabigatran CAL Set 4x1 mL

REF 5090210  
LOT 0691B00.01  
 2021-06-30

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Dabigatran Cal 1	1T82B01	2021-06-30
1x1mL	TECHNOVIEW Dabigatran Cal 2	1U82B01	2021-06-30
1x1mL	TECHNOVIEW Dabigatran Cal 3	1V82B01	2021-06-30
1x1mL	TECHNOVIEW Dabigatran Cal 4	1W82B01	2021-06-30

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Dabigatran Cont low LOT 3745B02	116.0 – 215.4 ng/mL	197.0 OK
	TECHNOVIEW Dabigatran Cont high LOT 3642B03	236.7 – 394.5 ng/mL	374.4 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

**Durchführung von  
Prepared by**

  
**Stefan STEINER**

28. NOV. 2019

Datum / Date:

**Freigabe von  
Approved by**

  
**Suzana GRESAK PODAKAR**

28. NOV. 2019

Datum / Date: