



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1000.01

TECHNOVIEW Apixaban CAL Set 5x1mL

REF 5090269

LOT 1091B00.02

🕒 2021-08-31

Inhalt Content	Reagenz Reagent	LOT	🕒
1x1mL	TECHNOVIEW Apixaban Cal 1	1191B00	2021-08-31
1x1mL	TECHNOVIEW Apixaban Cal 2	1291B00	2021-08-31
1x1mL	TECHNOVIEW Apixaban Cal 3	1391B00	2021-08-31
1x1mL	TECHNOVIEW Apixaban Cal 4	1491B00	2021-08-31
1x1mL	TECHNOVIEW Apixaban Cal 5	1591B00	2021-08-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Apixaban Cont Low LOT 3981B00	64.2 – 119.2 ng/mL	98.7 OK
	TECHNOVIEW Apixaban Cont High LOT 3D81B00	229.3 – 382.1 ng/mL	305.4 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	

Durchführung von
Prepared by

Eva Maria HAGER

23. MRZ. 2020

Datum / Date:

Freigabe von
Approved by

Suzana GRESAK PODAKAR

24. MRZ. 2020

Datum / Date: