
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3D00.01
---	---	-----------

TECHNOVIEW® Apixaban CONT High 5x1 mL

REF 5090270
LOT 3D01C00.01
 2022-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		214.7 – 357.9 ng/mL	286.3	OK
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban	214.7 – 357.9 ng/mL	282.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**


Stefan STEINER

27. APR. 2020
Datum / Date:

**Freigabe von
Approved by**


Eva Maria HAGER

27. APR. 2020
Datum / Date: