
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3900.01
---	---	-----------


**TECHNOVIEW® Apixaban CONT Low 5x1 ml**

REF	5090271
LOT	3901C00.01
	2022-02-28

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		76.6 – 142.2 ng/mL	109.4	OK
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	76.6 – 142.2 ng/mL	105.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

**Stefan STEINER**



24. APR. 2020

Datum / Date:

**Freigabe von**  
**Approved by**

**Eva Maria HAGER**



24. APR. 2020

Datum / Date: