

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1R00.01
---	---	-----------

Coagulation Reference, 5x1mL

REF 5220110
LOT 1R91B00.03
 2021-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.0 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	Fibrinogen, Clauss method, diluted plasma CCN LOT 1P02C00	2.57 – 3.85 g/L	3.57	OK
	Fibrinogen, Clauss method, diluted plasma CCA LOT 3P91B00	0.93 – 1.39 g/L	1.26	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Ladin NASUFI

04. JUNI 2020

Datum / Date:

Freigabe von
Approved by


Christoph ENNE

04. JUNI 2020

Datum / Date: