



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0Z00.02

TECHNOCHROM<sup>®</sup> FXIII 3 x 3 mL

REF 5360010  
LOT 0Z03C00.02  
🕒 2021-04-30

Inhalt	Reagenz	LOT	🕒
3 x 3 mL	FXIII Activator Reagent	2002B100/1	2021-07-31
3 x 3 mL	FXIII Detection Reagent	2002100/2	2021-04-30
3 x 3 mL	FXIII NADPH	2002100/3	2021-04-30
3 x 1 mL	FXIII Inhibitor Reagent	2002100/4	2021-04-30
1 x 6 mL	Stabilizer solution	2002100/5	2021-04-30

Test		Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P02C00	60.8 – 101.3 %	78.4
	Coagulation Control A Lot: 3P91B00	24.0 – 40.0 %	34.9

Durchführung von  
Prepared by

Michael KAFKA

28. JULI 2020

Datum / Date:

Freigabe von  
Approved by

Eva Maria HAGER

28. JULI 2020

Datum / Date: