



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ6PT00AZ

TECHNOZYM<sup>®</sup> ADAMTS-13 INH Calibrator Set 5x0.5mL

REF 5450461

LOT 6PT01C0.03

⌚ 2021-11-30

Inhalt Content	Reagenz Reagent	LOT	⌚
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT92B0	2021-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.56	OK
	Calibrator 5	< 0.1	0.01	OK
Verwendbarkeit Usability	High Control (RPY92B0)	56.36 – 93.94 U/mL	84.35	OK
	Low Control (RPZ92B0)	15.11 – 35.27 U/mL	27.14	OK

Durchführung von  
Prepared by

Fikret NASUFI *Fikret Nasuh*

28. JULI 2020

Datum / Date:

Freigabe von  
Approved by

Michael KAFKA *[Signature]*

29. JULI 2020

Datum / Date: