

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ...00AZ
-----------------------------------------------------------------------------------	-------------------------------------------------------------	-----------

TECHNOZYM[®] PAP Complex ELISA Kit 96T.

REF TC12060
LOT 120601C0.01
 2021-11-30

Inhalt Content	Reagenz Reagent	LOT	
1 x	PAP Plate	RGY92B0	2021-12-31
1 x 0.3mL	anti PAP POX	RBK92B0	2021-12-31
2 x 20mL	Wash buffer Concentrate	RBE01C0	2022-04-30
1 x 20mL	Dilution Buffer	RAF93B0	2021-11-30
1 x 12mL	TMB Substrate	RKN01C0	2022-01-31
1 x 12mL	Stop solution	RKK01C0	2022-01-31
1 x 0.5mL	PAP-Calibrator 1	1BM01C0	2022-05-31
1 x 0.5mL	PAP-Calibrator 2	2BM01C0	2022-05-31
1 x 0.5mL	PAP-Calibrator 3	3BM01C0	2022-05-31
1 x 0.5mL	PAP-Calibrator 4	4BM01C0	2022-05-31
1 x 0.5mL	PAP-Calibrator 5	5BM01C0	2022-05-31
1 x 0.5mL	PAP-high Control	ABM01C0	2022-05-31
1 x 0.5mL	PAP-low Control	BBM01C0	2022-05-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	OD: 1 – 2.5	1.81	OK
	Calibrator 5	OD: < 0.2	0.05	OK
Verwendbarkeit Usability	High Control (ABM01C0)	68.4 – 114.0 ng/mL	92.75	OK
	Low Control (BBM01C0)	28.8 – 67.2 ng/mL	47.02	OK

Durchführung von
Prepared by

Fikret NASUFI 

21. JULI 2020

Datum / Date:

Freigabe von
Approved by

Michael KAFKA 

21. JULI 2020

Datum / Date: