

## PL Chip Reservoir set for T-TAS<sup>®</sup> 01

### Package Insert

#### INTENDED USE:

The PL Chip Reservoir set for T-TAS 01 contains accessories for use with the PL Chip for T-TAS 01.

#### SUMMARY:

The PL Chip Reservoir set for T-TAS 01 contains disposable accessories that are used with the PL Chip for T-TAS 01 for measuring primary hemostatic ability of whole blood samples.

For additional details about the use of the PL Chip Reservoir set for T-TAS 01, please consult the PL Chip for T-TAS instructions for use and the T-TAS 01 User's Manual.

#### MATERIALS PROVIDED:

The PL Chip Reservoir set for T-TAS 01 contains 100 sets of reservoirs, caps, and over-caps.

Item	Contents	Catalog Number
PL Chip Reservoir set for T-TAS 01	100 sets	18003

#### WARNINGS AND PRECAUTIONS:

- For in vitro diagnostic use only.
- For professional use only.
- Used reservoirs, caps and over-caps are potentially infectious. Proper handling and disposal methods should be followed in accordance with local, state and federal regulations.
- Carefully follow the instructions and procedures described in this package insert.
- Do not use materials that are bent, deformed, broken, or contain foreign matter.










#### STORAGE AND HANDLING REQUIREMENTS:


Store unused PL Chip Reservoir set contents at room temperature. Do not remove the reservoir, caps, or over-caps from their bag until ready for use.

#### ASSISTANCE:

For assistance, please contact your local distributor.


**DEFINITION OF SYMBOLS:**

Symbol	Definition
	Do Not Re-use
	CE Mark
	Consult Instructions for use
	Batch Code
	Catalogue number
	<i>In Vitro</i> diagnostic medical device
	Manufacturer
	Authorized representative In the European Community
	Contains sufficient for <n> tests

Symbol	Definition
	This product is restricted to sale by or on the order of a licensed healthcare practitioner



FUJIMORI KOGYO CO., LTD.  
1-23-7 Nishi-Shinjuku,  
Shinjuku-ku, Tokyo 160-0023 Japan



Medical Device Safety Service GmbH  
Schiffgraben 41 30175 Hannover, Germany

T-TAS® is a registered trademark of FUJIMORI KOGYO CO., LTD.

©2019 FUJIMORI KOGYO CO., LTD.

All rights reserved

