



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ6L00.01

Att.06

6L91BA0.01

## Siron LIS (aPTT liquid) 10x4 ml

REF 5035119  
LOT 6L91BA0.01  
2021-12-31

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result		
visuelle Kontrolle visual inspection	Cap color	grey	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Appearance	clear liquid	confirmed	OK	
Kontrollwiederfindung aPTT am Ceveron® Alpha Recovery of controls	<b>Control N</b> LOT 1P02C00	25.2 - 34.0 sec	29.2 sec	OK
aPTT on Ceveron® Alpha	<b>Control A</b> LOT 3P91B00	66.6 - 90.0 sec	76.7 sec	OK

Durchgeführt von  
Prepared by

*Alessandro Benigni*  
Alessandro BENIGNI

*11-Sep-2020*

Datum / Date:

Freigabe von  
Approved by

*Suzana Gresak Podakar*  
Suzana GRESAK PODAKAR

*14. SEP. 2020*

Datum / Date: