



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0900.01

TECHNOVIEW® Rivaroxaban Cal Set 5x1 mL

REF 5090170
LOT 0901C00.01
⌚ 2022-05-31

Inhalt Content	Reagenz Reagent	LOT	⌚
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 1	5101C00	2022-05-31
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 2	5601C00	2022-05-31
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 3	5701C00	2022-05-31
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 4	5801C00	2022-05-31
1x1 mL	TECHNOVIEW® Rivaroxaban Cal 5	5901C00	2022-05-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability Anti-Xa Rivaroxaban Ceveron® alpha	Rivaroxaban Cont M LOT 3591B00	80.6 – 134.4 ng/mL	121.9	OK
	Rivaroxaban Cont L LOT 3401C00	40.3 – 74.8 ng/mL	59.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von
Prepared by

Eva Maria HAGER

11. SEP. 2020

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

14. SEP. 2020

Datum / Date: