



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ1L00.01

LOT 1L02C00.01

Platelet Poor Plasma 5x2 mL

REF 5343022

LOT 1L02C00.01

🕒 2022-07-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
pH	7.4 – 7.8	7.7	OK	
Restfeuchte Residual moisture	≤1%	0.41	OK	
OD _{320nm}	≤10	5.5	OK	
Verwendbarkeit Usability Ceveron® alpha	LA Ratio: Technoclot® LA Screen, LOT 6Y91B00 Technoclot® LA Confirm, LOT 6X91B00	0.83 – 1.38	1.12	OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK		

Durchführung von
Prepared by

Eva Maria HAGER

19. AUG. 2020
Datum / Date:

Freigabe von
Approved by

Stefan STEINER

19. AUG. 2020
Datum / Date: