



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ0J00.02

Att.14

0J03C00.01

## TECHNOCHROM C1-INH Reagent Kit

REF	5345003
LOT	0J03C00.01
	2022-04-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	1mL	Coagulation Control N for C1-INH	1P03CC1H	2022-06-30
1x	1mL	Coagulation Control A for C1-INH	3P01CC1H	2022-06-30
1x	see label	Coagulation Reference for C1-INH	1R01CC1H	2022-06-30
1x	3mL	Substrate f. C1 18 µmol	8C91B00	2023-09-30
1x	3mL	C1-Esterase	7B02C00	2022-04-30
1x	25mL	Probenpuffer A	9M01C00	2025-02-28
1x	20mL	Reaktionspuffer B	9N01C00	2025-02-28

Analytische Verfahren Analytical Procedures		Anforderung Requirement	Ergebnis Result	
visuelle Kontrolle visual inspection	Content	according to packaging SOP	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
Kontrollwiederfindung am Ceveron® alpha	<b>Control N</b> LOT 1P03CC1H	73.6   -   122.6   %	104.2   %	OK
Recovery of controls on Ceveron® alpha	<b>Control A</b> LOT 3P01CC1H	32.4   -   54.0   %	48.7   %	OK

Durchgeführt von  
Prepared by

**Eva Maria HAGER**

31. AUG. 2020

Datum / Date:

Freigabe von  
Approved by

**Stefan STEINER**

31. AUG. 2020

Datum / Date: