
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Z00.02
---	---	-----------

TECHNOCHROM[®] FXIII 3 x 3 mL

REF	5360010
LOT	0Z04C00.01
	2021-08-31

Inhalt	Reagenz	LOT	
3 x 3 mL	FXIII Activator Reagent	2003100/1	2021-08-31
3 x 3 mL	FXIII Detection Reagent	2003100/2	2021-08-31
3 x 3 mL	FXIII NADPH	2003100/3	2021-08-31
3 x 1 mL	FXIII Inhibitor Reagent	2003100/4	2021-08-31
1 x 6 mL	Stabilizer solution	2003100/5	2021-08-31

Test	Spezifikation	Ergebnis
Wiederfindung der Kontrollen / Recovery of controls	Coagulation Control N Lot: 1P02C00	60.8 – 101.3 % 78.4
	Coagulation Control A Lot: 3P91B00	24.0 – 40.0 % 33.8

Durchführung von
Prepared by

Michael KAFKA

28. AUG. 2020

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI *Fikret Nasufi*

28. AUG. 2020

Datum / Date: