

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQxUT0.02
		Att.21
		7UT02C0.01

TECHNOZYM® ADAMTS-13 Activity Control Set 2 x 0.5 mL

REF	5450763
LOT	7UT02C0.01
	2022-04-30

Inhalt Content	Volumen volume	Reagenz Reagent	LOT	
1x	0.5mL	ADAMTS-13 Activity High Control	RUY01C0	2022-04-30
1x	0.5mL	ADAMTS-13 Activity Low Control	RUZ01C0	2022-04-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Content	according to packaging SOP
	Vial label	according to packaging SOP
	Vial barcode	according to packaging SOP
	Box label	according to packaging SOP
	Box barcode	according to packaging SOP
	Package insert	present
	Batch table	present
Kontrollwiederfindung Recovery of controls	ADAMTS-13 Activity High Control LOT RUY01C0	0.52 - 0.86 IU/mL 0.77 IU/mL OK
Recovery of controls	ADAMTS-13 Activity Low Control LOT RUZ01C0	0.13 - 0.29 IU/mL 0.24 IU/mL OK

Durchgeführt von
Prepared by


Fikret NASUFI

03. JULI 2020

Datum / Date:

Freigabe von
Approved by


Eva Maria HAGER

03. JULI 2020

Datum / Date: