



Certificate of Analysis

REAADS
Monoclonal Free
Protein S Test Kit

Kit Lot #: MFPS-201
Plate Lot #: FSW-069
Product #: 051-001
Expiration Date: 2021-08-14

Quality Control Results

Testing Date: 2020-03-11
Manufacture Date: 2020-02-18

Absorbance	Value	Specification	Result by Assay
Reference Plasma 100% OD	87 % of Normal	≥ 1.0	1.207
Reagent Blank		≤ 0.250	0.150
Reference Curve R²		≥ 0.95	0.98947
Control Response	Label Range		Result by Assay
Normal Control	88 – 132 %		118.4 %
Deficient Control	14 – 46 %		31.7 %

This signed Certificate of Analysis affirms that the product has met the requirements of the relevant specifications as set by Corgenix, Inc.

This product is warranted to perform as described in the package insert. Corgenix, Inc. disclaims any implied warranty of merchantability or fitness for a particular use, and in no event shall Corgenix, Inc. be liable for consequential damage.

QA Approval By: Greg Hartvigsen

Signed:  2020-03-12

Title: Senior Quality Assurance Associate

Corgenix, Inc.
11575 Main Street, Suite 400
Broomfield, CO 80020
Phone: (303) 457-4345
Fax: (303) 457-4519