



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQB400AZ

Ceveron® TGA RC Low Kit 120T

REF 5006013

LOT B403C00.01

🕒 2021-05-31

Inhalt Content	Reagenz Reagent	LOT	🕒
2x1 mL	Ceveron® TGA RC Low	7R02C	2022-07-31
2x1.5 mL	Ceveron® TGA BUF	9Q91B	2021-11-30
2x3 mL	Ceveron® TGA SUB	8E91B00	2021-11-30
2x1 mL	Ceveron® TGA Control High	1Y02C00	2021-05-31
2x1 mL	Ceveron® TGA Control Low	2702C00	2021-05-31
1x25 mL	Ca-Chloridlösung 25 mmol/L	9C01CB0	2025-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Kontrollwiederfindung Recovery of controls	Ceveron® TGA Control-High (nM Peak Thrombin) LOT 1Y02C00	188 – 314	242.3	OK
	Ceveron® TGA Control-Low (nM Peak Thrombin) LOT 2702C00	87 – 145	110.7	OK
Etiketten und Barcode Überprüfung Label and bar code verification		OK		

Durchführung von
Prepared by

Michael KAFKA 

05. AUG. 2020

Datum / Date:

Freigabe von
Approved by

Fikret NASUFI 

05. AUG. 2020

Datum / Date: