



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ0U00.02

AK-Calibrant 4 x 1 mL

REF 5010004

LOT 0U02C00.02

2022-09-30

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	AK Calibrant A	1N84B02	2022-09-30
1x1mL	AK Calibrant B	3L88B02	2022-09-30
1x1mL	AK Calibrant C	3M8DB02	2022-10-31
1x1mL	AK Calibrant D	3N8EB02	2022-11-30

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P02C00 Technoplastin HIS	94.7 – 119.3 % 0.84 – 1.06 INR	107.0 0.99	OK OK
	CCA LOT 3P91B00 Technoplastin HIS	23.0 – 29.0 % 2.78 – 3.50 INR	25.3 3.21	OK OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Christoph ENNE

11. AUG. 2020

Datum / Date:

Freigabe von  
Approved by

Eva Maria HAGER

12. AUG. 2020

Datum / Date: