



# ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ6S0A.02

Att.06

6S01CA0.01

## Siron LS (aPTT liquid) 10x4 mL

REF

5035107

LOT

6S01CA0.01



2022-02-28

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result		
visuelle Kontrolle visual inspection	Cap color	grey	confirmed	OK
	Vial type	clear glass vial	confirmed	OK
	Vial label	according to packaging SOP	confirmed	OK
	Vial barcode	according to packaging SOP	confirmed	OK
	Box label	according to packaging SOP	confirmed	OK
	Box barcode	according to packaging SOP	confirmed	OK
	Package insert	present	confirmed	OK
	Appearance	clear liquid	confirmed	OK
Kontrollwiederfindung aPTT am Ceveron <sup>®</sup> Alpha Recovery of controls aPTT on Ceveron <sup>®</sup> Alpha	Control N LOT 1P71B01	21.8 - 29.6 sec	28.5 sec	OK
	Control A LOT 3P71B01	44.5 - 60.1 sec	58.4 sec	OK

Durchgeführt von  
Prepared by

Suzana GRESAK PODAKAR

30. MRZ. 2020

Datum / Date:

Freigabe von  
Approved by

Christoph ENNE

30. MRZ. 2020

Datum / Date: