



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ4K00.03

Att.03

4K06C0D.01

Fitzgerald Trait Plasma 2x1 ml

REF	5204006
LOT	4K06C0D.01
	2022-09-30

Analytische Verfahren Analytical Procedures	Anforderung Requirement	Ergebnis Result
visuelle Kontrolle visual inspection	Cap color	blue confirmed OK
	Vial type	clear glass vial confirmed OK
	Vial label	according to packaging SOP confirmed OK
	Vial barcode	according to packaging SOP confirmed OK
	Box label	according to packaging SOP confirmed OK
	Box barcode	according to packaging SOP confirmed OK
	Package insert	present confirmed OK
	Appearance before reconstitution	buff colored plug of lyophilized material confirmed OK
Appearance after reconstitution	buff colored liquid confirmed OK	
Virologie / Virology The testing methods applied were FDA-approved or CE marked.	HBsAg / HBsAg HIV-Ak / HIV-Ab HCV-AK / HCV-Ab Negative Negative Negative	Negative Negative Negative OK
Kontrollwiederfindung HMWK am Ceveron [®] alpha Recovery of controls	Control N LOT 1P02C00 Control A LOT 3P91B00	60.0 - 100.0 % 84.6 % OK 19.5 - 32.5 % 26.7 % OK

Durchgeführt von
Prepared by

Christoph ENNE

21. SEP. 2020

Datum / Date:

Freigabe von
Approved by

Stefan STEINER

22. SEP. 2020

Datum / Date: