



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ9C00.02

LOT 9C01CB0

Ca-Chloridlösung 25mmol/L, 25mL

REF 5277018  
LOT 9C01CB0  
2020-01-20  
2025-01-31

Analytische Verfahren Analytical Procedure		Anforderung Requirement	Ergebnis Result		
Volumen volume		≥ 25 mL	26	OK	
Sterilität sterility		0 CFU/mL	0	OK	
CaCl <sub>2</sub> -Gehalt CaCl <sub>2</sub> -Content		23.5 – 26.5 mmol/L	25.2	OK	
Kontrollwiederfindung Recovery of controls	aPTT, sec LOT 6S82BB0	CCN LOT 1P92B00	22.2 – 27.4 sec	26.6	OK
		CCA LOT 3P71B01	46.8 – 57.8 sec	52.1	OK

Durchführung von  
Prepared by

Eva Maria HAGER

04. FEB. 2020  
Datum / Date:

Freigabe von  
Approved by

Stefan STEINER

05. FEB. 2020  
Datum / Date: