

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ9J0A.01
---	---	-----------

Imidazol Puffer 25mL

REF 5410008
LOT 9J54BC1.01
 2025-01-31

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.2 – 7.6	7.5	OK
Verwendbarkeit Usability Ceveron® alpha Fibrinogen, Clauss Method	TCN LOT 1A99B0W	2.88 – 4.32 g/L	3.67 OK
	TCA LOT 3A91B00	0.94 – 1.40 g/L	1.27 OK
Etiketten Überprüfung Label Verification		OK	

**Durchführung von
Prepared by**



Lavdrim NASUFI

13. MAI 2020

Datum / Date:

**Freigabe von
Approved by**



Christoph ENNE

14. MAI 2020

Datum / Date: