
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQRP00AZ
---	---	----------

TECHNOZYM® ADAMTS-13 INH ELISA Kit 96T

REF 5450401
LOT RP0AC00.01
 2021-11-30

Inhalt Content	Reagenz Reagent	LOT	
1 x 12	Adamts13 INH Plate	RPS06C0	2022-08-31
1 x 0.3mL	Adamts13 INH Conjugate	RPV92B0	2021-11-30
1x 80mL	Wash buffer Concentrate	RNA02C0	2022-07-31
1 x 90mL	Incubation buffer	RNB02C0	2022-06-30
1 x 12mL	TMB Substrate	RKN01C0	2022-01-31
1 x 12mL	Stop solution	RKK02C0	2022-06-30
1 x 0.5mL	Adamts13 INH Calibrator 1	1PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 2	2PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 3	3PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 4	4PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Calibrator 5	5PT92B0	2021-11-30
1 x 0.5mL	Adamts13 INH High Control	RPY92B0	2021-11-30
1 x 0.5mL	Adamts13 INH Low Control	RPZ92B0	2021-12-31

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
OD	Calibrator 1	> 1.0	1.41	OK
	Calibrator 5	< 0.1	0.04	OK
Verwendbarkeit Usability	High Control (RPY92B0)	56.36 - 93.94 U/mL	79.65	OK
	Low Control (RPZ92B0)	15.11 - 35.27 U/mL	24.77	OK

Durchführung von
Prepared by

Fikret NASUFI *Fikret Nasufi*

26. SEP. 2020
Datum / Date:

Freigabe von
Approved by

Christoph Enne
Christoph ENNE

27. SEP. 2020
Datum / Date: