

CERTIFICATE OF ANALYSIS



PRODUCT NAME :	COAMATIC FACTOR VIII	EXPIRATION DATE:	2022-05
PRODUCT NUMBER :	82258563		
LOT NUMBER:	N0604819		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	PASS/FAIL
	MIN	MAX			
HIGH RANGE					
Slope	0.33	0.57	--	0.42	PASS
Intercept	---	0.12	--	0.11	PASS
r	0.990	---	--	0.998	PASS
LOW RANGE					
Slope	2.8	6.9	--	5.1	PASS
Intercept	---	0.21	--	0.14	PASS
r	0.990	---	--	0.999	PASS

KIT COMPONENTS

NAME	LOT NUMBER	VOLUME
COAMATIC FVIII FACTOR REAGENT	N0504637	3 mL
COAMATIC FVIII S-2765 + I-2581	N0604817	6 mL
COAMATIC FVIII BUFFER	N0604812	24 mL

have been manufactured and tested according to the requirements of their Manufacturing and Quality Control procedures and have been found to meet all

PREPARED BY: Nehal Patel *N.M. Patel* DATE: 6-19-20
 PRINT SIGNATURE 6/19/2020

TITLE: Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 6/19/20
 PRINT SIGNATURE 6/19/2020

TITLE: Technical Supervisor