

CERTIFICATE OF ANALYSIS

CHROMOGENIX


**Instrumentation
Laboratory**
 A Werfen Company

PRODUCT NAME:	CoaTest SP Factor FVIII Kit	EXPIRATION DATE:	2022-06
PRODUCT NUMBER:	82408663		
LOT NUMBER :	N0605414		

TEST DESCRIPTION	SPECIFICATIONS		UNITS	RESULTS	DISPOSITION
	MIN	MAX			
High Range					
Slope	0.0077	0.0112	--	0.0084	PASS
Intercept	--	0.250	--	0.128	PASS
Standard Curve (r)	0.99	--	--	0.998	PASS
Low Range					
Slope	0.020	0.060	--	0.036	PASS
Intercept	--	0.406	--	0.169	PASS
Standard Curve (r)	0.99	--	--	0.997	PASS

KIT COMPONENTS:

NAME	LOT NUMBER	VOLUME
COATEST SP FVIII CaCl ₂	N0605149	6 mL
COATEST SP FVIII Factor IXa + Factor X	N0605152	10 mL
COATEST SP FVIII S-2765 + I-2581	N0605310	12 mL
COATEST SP FVIII Buffer, stock solution	N0605151	20 mL
COATEST SP FVIII Phospholipid	N0605313	2 mL

PREPARED BY: Nehal Patel *N. m. Patel* DATE: 09/02/20
 PRINT SIGNATURE

TITLE Senior QC Analyst

APPROVAL SIGNATURE: Debra Lee Harz *Debra Lee Harz* DATE: 09/02/20
 PRINT SIGNATURE

TITLE Technical Supervisor