




Certificate of Analysis
M65 EpiDeath[®] ELISA


REF 10040

M65 EpiDeath[®] ELISA

REF 10040

Kit Lot No. LOT 10040-302	Expiry  2021-10
--	--

Component		A ₄₅₀ nm	LOT	
M65 EpiDeath Standard A	0 U/L	0,050	10040SA- 302	2022-01
M65 EpiDeath Standard B	200 U/L	0,144	10040SB- 302	2022-01
M65 EpiDeath Standard C	400 U/L	0,209	10040SC- 302	2022-01
M65 EpiDeath Standard D	800 U/L	0,435	10040SD- 302	2022-01
M65 EpiDeath Standard E	1200 U/L	0,655	10040SE- 302	2022-01
M65 EpiDeath Standard F	2000 U/L	1,056	10040SF- 302	2022-01
M65 EpiDeath Standard G	3000 U/L	1,588	10040SG- 302	2022-01
M65 EpiDeath Standard H	5000 U/L	2,612	10040SH- 302	2022-01
M5 Coated Microstrips			M5- 008	2023-02
M65 EpiDeath HRP Conjugate			10040C- 302	2022-04
M65 EpiDeath Conjugate Dilution Buffer			10040CD- 302	2022-05
TMB Substrate			VLVT- 003	2022-12
Stop Solution			VLVSS- 008	2023-02
Wash Tablet			VLVWT- 004	2021-10
Controls	Obtained Value (U/L)	Acceptable Range (U/L)	LOT	
M65 EpiDeath Control Low	270	±60 U/L	10040CL- 302	2022-01
M65 EpiDeath Control High	2583	±500 U/L	10040CH- 302	2022-01

Test date: 10.07.2020	Sign: 
-----------------------	---

