

Certificate of Analysis
M30 Apoptosense[®] ELISA

REF 10011






VLVbio AB, Hästholmsvägen 32, 131 30 Nacka, Sweden


Phone: +46 8 122 053 00 • www.vlvbio.com • info@vlvbio.com

M30 Apoptosense® ELISA

REF 10011

Kit Lot No. LOT 10011- 007	Expiry  2020-10
---	--

Component		A _{450 nm}	LOT		
M30 Standard A	0 U/L	0,105	10011SA-	007	2020-12
M30 Standard B	75 U/L	0,238	10011SB-	007	2020-12
M30 Standard C	150 U/L	0,378	10011SC-	007	2020-12
M30 Standard D	250 U/L	0,686	10011SD-	007	2020-12
M30 Standard E	500 U/L	1,528	10011SE-	007	2020-12
M30 Standard F	750 U/L	2,317	10011SF-	007	2020-12
M30 Standard G	1000 U/L	2,834	10011SG-	007	2020-12
M5 Coated Microstrips			M5-	007	2023-01
M30 HRP Conjugate			10011C-	007A	2020-11
M30 Conjugate Dilution Buffer			10011CD-	007	2020-10
TMB Substrate			VLVT-	004	2023-03
Stop Solution			VLVSS-	007	2022-11
Wash Tablet			VLVWT-	004	2021-10
Controls	Obtained Value (U/L)	Acceptable Range (U/L)	LOT		
M30 Control Low	120	±25 U/L	10011CL-	007	2020-12
M30 Control High	651	±100 U/L	10011CH-	007	2020-12

Test date: 24.03.2020	Sign: 
-----------------------	---

PEVIVA Products

ASSAYS

M30 Apoptosense[®] ELISA
Prod. No. 10011

M65[®] ELISA
Prod. No. 10020

M65 EpiDeath[®] ELISA
Prod. No. 10040

M30 CytoDeath[™] ELISA
Prod. No. 10900

ANTIBODIES

M30 CytoDEATH[™]

- **Unconjugated** **Prod. No. 10700**
- **Biotin** **Prod. No. 10750**
- **Fluorescein** **Prod. No. 10800**
- **Orange** **Prod. No. 10830**

M5 Keratin 18
Prod.No. 10600

M6 Keratin 18
Prod. No. 10650