

**Certificate of Analysis**  
**M30 Apoptosense<sup>®</sup> ELISA**



**REF** 10011




# M30 Apoptosense<sup>®</sup> ELISA

**REF** 10011

Kit Lot No. <b>LOT</b> 10011- 008	Expiry  2021-05
-----------------------------------------	--------------------------------------------------------------------------------------------------------

Component		A <sub>450 nm</sub>	LOT		
M30 Standard A	0 U/L	0,130	10011SA-	008	2021-06
M30 Standard B	75 U/L	0,250	10011SB-	008	2021-06
M30 Standard C	150 U/L	0,364	10011SC-	008	2021-06
M30 Standard D	250 U/L	0,631	10011SD-	008	2021-06
M30 Standard E	500 U/L	1,320	10011SE-	008	2021-06
M30 Standard F	750 U/L	2,244	10011SF-	008	2021-06
M30 Standard G	1000 U/L	2,836	10011SG-	008	2021-06
M5 Coated Microstrips			M5-	009	2023-06
M30 HRP Conjugate			10011C-	008	2021-05
M30 Conjugate Dilution Buffer			10011CD-	008	2021-05
TMB Substrate			VLVT-	006	2023-11
Stop Solution			VLVSS-	008	2023-02
Wash Tablet			VLVWT-	005	2023-01
Controls	Expected value (U/L)	Obtained value (U/L)	LOT		
M30 Control Low	125 ± 25 U/L	111,2U/L	10011CL-	008	2021-06
M30 Control High	650 ± 100 U/L	652,5U/L	10011CH-	008	2021-06

Test date: 11.09.2020	Sign: 
-----------------------	-------------------------------------------------------------------------------------------

