

Certificate of Analysis

M65 ELISA[®]

REF 10020





VLVbio AB, Hästholmsvägen 32, 131 30 Nacka, Sweden


Phone: +46 8 122 053 00 • www.vlvbio.com • info@vlvbio.com

M65[®] ELISA

REF 10020

Kit Lot No. LOT 10020-203A	Expiry  2021-03
---	--

Component		A ₄₅₀ nm	LOT		
M65 Standard A	0 U/L	0,051	10020SA-	203	2021-03
M65 Standard B	125 U/L	0,188	10020SB-	203	2021-03
M65 Standard C	250 U/L	0,345	10020SC-	203	2021-03
M65 Standard D	500 U/L	0,703	10020SD-	203	2021-03
M65 Standard E	750 U/L	1,042	10020SE-	203	2021-03
M65 Standard F	1200 U/L	1,632	10020SF-	203	2021-03
M65 Standard G	2000 U/L	2,545	10020SG-	203	2021-03
M6 Coated Microstrips			M6-	003	2023-02
M65 HRP Conjugate			10020C-	203	2021-07
M65 Conjugate Dilution Buffer			10020CD-	203	2021-11
TMB Substrate			VLVT-	004	2023-03
Stop Solution			VLVSS-	008	2023-02
Wash Tablet			VLVWT-	004	2021-10
Controls	Expected value (U/L)	Obtained value (U/L)	LOT		
M65 Control Low	160 – 240	226	10020CL-	203	2021-03
M65 Control High	850 – 1 150	1062	10020CH-	203	2021-03

Test date: 03.04.2020	Sign: 
-----------------------	---

PEVIVA Products

ASSAYS

M30 Apoptosense[®] ELISA
Prod. No. 10011

M65[®] ELISA
Prod. No. 10020

M65 EpiDeath[®] ELISA
Prod. No. 10040

M30 CytoDeath[™] ELISA
Prod. No. 10900

ANTIBODIES

M30 CytoDEATH[™]

- **Unconjugated** **Prod. No. 10700**
- **Biotin** **Prod. No. 10750**
- **Fluorescein** **Prod. No. 10800**
- **Orange** **Prod. No. 10830**

M5 Keratin 18
Prod.No. 10600

M6 Keratin 18
Prod. No. 10650